



**CALIBRATION CERTIFICATE**

**Type:** X-ray and gamma radiation dosimeter AT1123

Date of calibration: 15.06.2022.

**S/N:** 56557

Measurement limits:

**AT1123:  $\gamma$  50 nSv/h – 10 Sv/h;**

Measurement error:

**AT1123:  $\pm 15\%$**

Operating conditions:

- Air temperature +19,8 °C
- Atmospheric pressure 98,5 kPa
- Relative humidity 72,2 %
- Gamma radiation background 94,2 nSv/h

The instrument is calibrated on; standard dosimetry facility AT-110, N 013 reg. N 40425-09, the Certificate of Compliance N C-B/15-11-2021/112536221 on 15.11.2021 issued by FGUP «D.I.Mendeleyev VNIIM», St. Petersburg, Russia); standard dosimetry facility AT-130, N 015, reg. N 44761-10, the Certificate of Compliance C-B/15-11-2021/112536213 on 15.11.2021 issued by FGUP «D.I.Mendeleyev VNIIM», St. Petersburg, Russia);

**Calibration data**

**AT1123 ( $\gamma$ ) s/n: 56557**

Dose rate at check point $\dot{H}_{oi}(10)$	Radiation source number	Distance to source, $R, cm$	Dose rate measurement at check point,				Relative gamma radiation dose rate measurement error $\theta_{mpi}, \%$	Confidence limit of the intrinsic relative error $\Delta_i, \%$ during calibration	Limits of intrinsic relative error, % not above	
			Back-ground, $nSv/h$	Measured value $\dot{H}_i^*(10)$						Average value, $H_i(10)$
				$H_1$	$H_2$	$H_3$				
0,07 $\mu Sv/h$	263	194.5	94,2	0,070	0,068	0,072	0,070	0,00	5,50	$\pm 15$
0,7 $\mu Sv/h$	0HA	228.6	94,2	0,71	0,68	0,70	0,70	0,00	5,50	
7 $\mu Sv/h$	0HA	73.4	—	7,13	7,02	6,85	7,00	0,00	5,50	
70 $\mu Sv/h$	9XK	158.8	—	69,4	70,0	67,6	69,0	-1,43	5,61	
0,7 mSv/h	9XK	51,0	—	0,68	0,69	0,73	0,70	0,00	5,50	
7,0 mSv/h	043	343.4	—	7,17	6,90	6,92	7,00	0,00	4,40	
70 mSv/h	043	110.3	—	71,6	71,9	66,4	70,0	0,00	4,40	
0,7 Sv/h	163	227.6	—	0,69	0,71	0,69	0,70	0,00	4,40	
7 Sv/h	163	73.3	—	7,02	6,93	6,75	6,90	-1,43	4,54	

Calibrated by:

V. Pisarenko

\_\_\_\_\_  
(signature)

Technical control:

N. Kurbatova

\_\_\_\_\_  
(signature)