



Polimaster Ltd.

Legal address: 112-3n, room 53, M. Bogdanovich str., Minsk, 220040

Mailing address:

51, F. Skorina str., Minsk, 220141

Republic of Belarus

Phone: + 375 17 268 6819

Fax: + 375 17 264 2356

www.polimaster.com

E-mail: polimaster@polimaster.com

Calibration protocol

Nº 1-BY/04/2022 dated April 26, 2022

Date when calibrated

from April 5, 2022 till April 15, 2022.

Item calibrated

Dosimetric gamma-radiation facility UDG-PM9000 №002.

Customer

UAB Polimaster Europe, Ežero g. 4, Didžiasalio k., Nemėžio sen. LT-13264 Vilnius, Lietuva.

Organization, that performed the calibration

Polimaster Ltd., 112-3n, room 53, Bogdanovicha str., Minsk, 220040, Belarus.

Method of calibration

Calibration of UDG-PM9000 facility gamma-radiation fields is performed in terms of air kerma using the direct measurements method. Calibration is executed according to own draft calibration procedure.

Calibration is performed by using

PTW series electrometer UNIDOS^{webline} type T10021 № 000759 with cavity ionization chambers:

- TM32005 (TK-30) № 000097;
- TM32002 (LS-01) № 000498;
- TM32003 (LS-10) № 000163.

The above measuring equipment has established degrees of equivalence to the reference values of GET 8-2011 National Standard of Russian Federation confirmed by a calibration certificate.

Calibration results including uncertainty

Results of calibration (including expanded uncertainty) are presented in tables (1 - 7) in this calibration certificate. Ambient dose equivalent rates $\dot{H}^*(10)$ and personal dose equivalent rates $\dot{H}_p(10)$ are also given in the tables. These values are received using multiplication of measured air kerma rate values by the conversion coefficients $h^*(10)_K$ and $h_p(10)_K$. The conversion coefficients for radionuclides ^{241}Am , ^{137}Cs and ^{60}Co are given in the Annex «A» to this calibration protocol.



Continuation of the calibration protocol № 1-BY/04/2022 dated April 26, 2022

Additional information

Fields sizes of UDG-RM9000 calibration facility, recommended half-life periods for radionuclide sources and conversation coefficients to operational quantities are given in the Annex «A» to this calibration protocol.

Geometric conditions during calibration

Calibration of UDG-PM9000 calibration facility under conditions of collimated gamma beams is carried out in terms of air kerma on the axis of gamma radiation beam at distances from 0.5 m to 6.5 m measured from the center of the active part of the source positioned in the collimator. Distance adjustment was carried out using UDG-PM9000 facility scale.

Environmental conditions during calibration

- air temperature from 19.6 °C to 20.8 °C;
- air pressure from 96.65 to 100.74 kPa;
- relative humidity from 40.4 % to 47.8 %.

Signature of the person who has performed the calibration



(signature)

Valiantsin Zubarau
(decryption of signature)

Continuation of the calibration protocol № 1-BY/04/2022 dated April 26, 2022

1. UDG-PM9000 facility air kerma rates

1.1. Air kerma rate, ambient dose equivalent rate and personal dose equivalent rate of gamma radiation from ^{137}Cs source of type ИГИ-Ц-8-2 № 150 (collimator № 2), the results of measurements are dated April 20, 2022

Table 1

Distance from center of active part of a source to operating point R, m	Air kerma rate \dot{K}_a , Gy/s	Expanded uncertainty $U(\dot{K}_a)$, %	Ambient dose equivalent rate $\dot{H}^*(10)$, Sv/s	Expanded uncertainty $U(\dot{H}^*(10))$, %	Personal dose equivalent rate $\dot{H}_p(10)$, Sv/s	Expanded uncertainty $U(\dot{H}_p(10))$, %
0.5	$2.671 \cdot 10^{-3}$	3.0	$3.232 \cdot 10^{-3}$	3.6	$3.232 \cdot 10^{-3}$	3.6
0.7	$1.325 \cdot 10^{-3}$	3.0	$1.604 \cdot 10^{-3}$	3.6	$1.604 \cdot 10^{-3}$	3.6
1.0	$6.376 \cdot 10^{-4}$	3.0	$7.715 \cdot 10^{-4}$	3.6	$7.715 \cdot 10^{-4}$	3.6
1.4	$3.214 \cdot 10^{-4}$	3.0	$3.889 \cdot 10^{-4}$	3.6	$3.889 \cdot 10^{-4}$	3.6
2.0	$1.551 \cdot 10^{-4}$	3.0	$1.877 \cdot 10^{-4}$	3.6	$1.877 \cdot 10^{-4}$	3.6
3.0	$6.788 \cdot 10^{-5}$	3.0	$8.214 \cdot 10^{-5}$	3.6	$8.214 \cdot 10^{-5}$	3.6
4.0	$3.789 \cdot 10^{-5}$	3.0	$4.584 \cdot 10^{-5}$	3.6	$4.584 \cdot 10^{-5}$	3.6
5.0	$2.411 \cdot 10^{-5}$	3.0	$2.918 \cdot 10^{-5}$	3.6	$2.918 \cdot 10^{-5}$	3.6
6.0	$1.670 \cdot 10^{-5}$	3.0	$2.020 \cdot 10^{-5}$	3.6	$2.020 \cdot 10^{-5}$	3.6
6.5	$1.422 \cdot 10^{-5}$	3.0	$1.721 \cdot 10^{-5}$	3.6	$1.721 \cdot 10^{-5}$	3.6

1.2. Air kerma rate, ambient dose equivalent rate and personal dose equivalent rate of gamma radiation from ^{137}Cs source of type ИГИ-Ц-4-6 № 8Х4 (collimator № 1), the results of measurements are dated April 20, 2022

Table 2

Distance from center of active part of a source to operating point R, m	Air kerma rate \dot{K}_a , Gy/s	Expanded uncertainty $U(\dot{K}_a)$, %	Ambient dose equivalent rate $\dot{H}^*(10)$, Sv/s	Expanded uncertainty $U(\dot{H}^*(10))$, %	Personal dose equivalent rate $\dot{H}_p(10)$, Sv/s	Expanded uncertainty $U(\dot{H}_p(10))$, %
0.5	$1.111 \cdot 10^{-5}$	3.0	$1.344 \cdot 10^{-5}$	3.6	$1.344 \cdot 10^{-5}$	3.6
0.7	$5.586 \cdot 10^{-6}$	3.0	$6.759 \cdot 10^{-6}$	3.6	$6.759 \cdot 10^{-6}$	3.6
1.0	$2.706 \cdot 10^{-6}$	3.0	$3.274 \cdot 10^{-6}$	3.6	$3.274 \cdot 10^{-6}$	3.6
1.4	$1.369 \cdot 10^{-6}$	3.0	$1.656 \cdot 10^{-6}$	3.6	$1.656 \cdot 10^{-6}$	3.6
2.0	$6.636 \cdot 10^{-7}$	3.0	$8.029 \cdot 10^{-7}$	3.6	$8.029 \cdot 10^{-7}$	3.6
3.0	$2.901 \cdot 10^{-7}$	3.1	$3.511 \cdot 10^{-7}$	3.7	$3.511 \cdot 10^{-7}$	3.7
4.0	$1.618 \cdot 10^{-7}$	3.1	$1.958 \cdot 10^{-7}$	3.7	$1.958 \cdot 10^{-7}$	3.7
5.0	$1.031 \cdot 10^{-7}$	3.1	$1.247 \cdot 10^{-7}$	3.7	$1.247 \cdot 10^{-7}$	3.7
6.0	$7.119 \cdot 10^{-8}$	3.1	$8.614 \cdot 10^{-8}$	3.7	$8.614 \cdot 10^{-8}$	3.7
6.5	$6.066 \cdot 10^{-8}$	3.1	$7.340 \cdot 10^{-8}$	3.7	$7.340 \cdot 10^{-8}$	3.7

Signature of the person who has performed the calibration

Valiantsin Zubaran

(decryption of signature)

Continuation of the calibration protocol № 1-BY/04/2022 dated April 26, 2022

1.3. Air kerma rate, ambient dose equivalent rate and personal dose equivalent rate of gamma radiation from ^{137}Cs source of type ИГИ-Ц-3-10 № 054 (collimator № 3, cell № 2), the results of measurements are dated April 20, 2022

Table 3

Distance from center of active part of a source to operating point R, m	Air kerma rate \dot{K}_a , Gy/s	Expanded uncertainty $U(\dot{K}_a)$, %	Ambient dose equivalent rate $\dot{H}^*(10)$, Sv/s	Expanded uncertainty $U(\dot{H}^*(10))$, %	Personal dose equivalent rate $\dot{H}_p(10)$, Sv/s	Expanded uncertainty $U(\dot{H}_p(10))$, %
0.5	$1.689 \cdot 10^{-7}$	3.1	$2,044 \cdot 10^{-7}$	3.7	$2,044 \cdot 10^{-7}$	3.7
0.7	$8,460 \cdot 10^{-8}$	3.1	$1,024 \cdot 10^{-7}$	3.7	$1,024 \cdot 10^{-7}$	3.7
1.0	$4,082 \cdot 10^{-8}$	3.1	$4,939 \cdot 10^{-8}$	3.7	$4,939 \cdot 10^{-8}$	3.7
1.4	$2,048 \cdot 10^{-8}$	3.1	$2,478 \cdot 10^{-8}$	3.7	$2,478 \cdot 10^{-8}$	3.7
2.0	$9,964 \cdot 10^{-9}$	3.1	$1,206 \cdot 10^{-8}$	3.7	$1,206 \cdot 10^{-8}$	3.7
3.0	$4,373 \cdot 10^{-9}$	3.1	$5,291 \cdot 10^{-9}$	3.7	$5,291 \cdot 10^{-9}$	3.7
4.0	$2,447 \cdot 10^{-9}$	3.1	$2,961 \cdot 10^{-9}$	3.7	$2,961 \cdot 10^{-9}$	3.7
5.0	$1,553 \cdot 10^{-9}$	3.1	$1,879 \cdot 10^{-9}$	3.7	$1,879 \cdot 10^{-9}$	3.7
6.0	$1,088 \cdot 10^{-9}$	3.1	$1,316 \cdot 10^{-9}$	3.7	$1,316 \cdot 10^{-9}$	3.7

1.4. Air kerma rate, ambient dose equivalent rate and personal dose equivalent rate of gamma radiation from ^{137}Cs source of type ИГИ-Ц-3-8 № 4CA (collimator № 3, cell № 1), the results of measurements are dated April 20, 2022

Table 4

Distance from center of active part of a source to operating point R, m	Air kerma rate \dot{K}_a , Gy/s	Expanded uncertainty $U(\dot{K}_a)$, %	Ambient dose equivalent rate $\dot{H}^*(10)$, Sv/s	Expanded uncertainty $U(\dot{H}^*(10))$, %	Personal dose equivalent rate $\dot{H}_p(10)$, Sv/s	Expanded uncertainty $U(\dot{H}_p(10))$, %
0.5	$3,647 \cdot 10^{-8}$	3.1	$4,413 \cdot 10^{-8}$	3.7	$4,413 \cdot 10^{-8}$	3.7
0.7	$1,828 \cdot 10^{-8}$	3.1	$2,212 \cdot 10^{-8}$	3.7	$2,212 \cdot 10^{-8}$	3.7
1.0	$8,833 \cdot 10^{-9}$	3.1	$1,069 \cdot 10^{-8}$	3.7	$1,069 \cdot 10^{-8}$	3.7
1.4	$4,462 \cdot 10^{-9}$	3.1	$5,399 \cdot 10^{-9}$	3.7	$5,399 \cdot 10^{-9}$	3.7
2.0	$2,173 \cdot 10^{-9}$	3.1	$2,629 \cdot 10^{-9}$	3.7	$2,629 \cdot 10^{-9}$	3.7
3.0	$9,365 \cdot 10^{-10}$	3.1	$1,133 \cdot 10^{-9}$	3.7	$1,133 \cdot 10^{-9}$	3.7
4.0	$5,319 \cdot 10^{-10}$	3.1	$6,435 \cdot 10^{-10}$	3.7	$6,435 \cdot 10^{-10}$	3.7
5.0	$3,348 \cdot 10^{-10}$	3.1	$4,051 \cdot 10^{-10}$	3.7	$4,051 \cdot 10^{-10}$	3.7

Signature of the person who has performed the calibration

(signature)

Valiantsin Zubara
(decryption of signature)

Continuation of the calibration protocol № 1-BY/04/2022 dated April 26, 2022

1.5. Air kerma rate, ambient dose equivalent rate and personal dose equivalent rate of gamma radiation from ^{137}Cs source of type Cs7.P03 № 0159/09 (collimator № 3, cell № 5), the results of measurements are dated April 20, 2022

Table 5

Distance from center of active part of a source to operating point R, m	Air kerma rate \dot{K}_a , Gy/s	Expanded uncertainty $U(\dot{K}_a)$, %	Ambient dose equivalent rate $\dot{H}^*(10)$, Sv/s	Expanded uncertainty $U(\dot{H}^*(10))$, %	Personal dose equivalent rate $\dot{H}_p(10)$, Sv/s	Expanded uncertainty $U(\dot{H}_p(10))$, %
0.5	$7.838 \cdot 10^{-9}$	3.1	$9.484 \cdot 10^{-9}$	3.7	$9.484 \cdot 10^{-9}$	3.7
0.7	$3.964 \cdot 10^{-9}$	3.1	$4.796 \cdot 10^{-9}$	3.7	$4.796 \cdot 10^{-9}$	3.7
1.0	$1.912 \cdot 10^{-9}$	3.1	$2.313 \cdot 10^{-9}$	3.7	$2.313 \cdot 10^{-9}$	3.7
1.4	$9.553 \cdot 10^{-10}$	3.1	$1.156 \cdot 10^{-9}$	3.7	$1.156 \cdot 10^{-9}$	3.7
2.0	$4.640 \cdot 10^{-10}$	3.3	$5.614 \cdot 10^{-10}$	3.8	$5.614 \cdot 10^{-10}$	3.8
3.0	$2.018 \cdot 10^{-10}$	3.3	$2.441 \cdot 10^{-10}$	3.9	$2.441 \cdot 10^{-10}$	3.9

1.6. Air kerma rate, ambient dose equivalent rate and personal dose equivalent rate of gamma radiation from ^{60}Co source of type ГИК-2-8 № T68 (collimator № 3, cell № 4), the results of measurements are dated April 20, 2022

Table 6

Distance from center of active part of a source to operating point R, m	Air kerma rate \dot{K}_a , Gy/s	Expanded uncertainty $U(\dot{K}_a)$, %	Ambient dose equivalent rate $\dot{H}^*(10)$, Sv/s	Expanded uncertainty $U(\dot{H}^*(10))$, %	Personal dose equivalent rate $\dot{H}_p(10)$, Sv/s	Expanded uncertainty $U(\dot{H}_p(10))$, %
0.5	$2.660 \cdot 10^{-8}$	3.1	$3.085 \cdot 10^{-8}$	3.7	$3.059 \cdot 10^{-8}$	3.7
0.7	$1.325 \cdot 10^{-8}$	3.1	$1.537 \cdot 10^{-8}$	3.7	$1.524 \cdot 10^{-8}$	3.7
1.0	$6.395 \cdot 10^{-9}$	3.1	$7.418 \cdot 10^{-9}$	3.7	$7.354 \cdot 10^{-9}$	3.7
1.4	$3.260 \cdot 10^{-9}$	3.1	$3.782 \cdot 10^{-9}$	3.7	$3.749 \cdot 10^{-9}$	3.7
2.0	$1.573 \cdot 10^{-9}$	3.2	$1.825 \cdot 10^{-9}$	3.8	$1.809 \cdot 10^{-9}$	3.8
3.0	$6.861 \cdot 10^{-10}$	3.2	$7.959 \cdot 10^{-10}$	3.8	$7.891 \cdot 10^{-10}$	3.8

1.7. Air kerma rate, ambient dose equivalent rate and personal dose equivalent rate of gamma radiation from ^{241}Am source of type ИГИА-3М № 179 (collimator № 3, cell № 3), the results of measurements are dated April 20, 2022

Table 7

Distance from center of active part of a source to operating point R, m	Air kerma rate \dot{K}_a , Gy/s	Expanded uncertainty $U(\dot{K}_a)$, %	Ambient dose equivalent rate $\dot{H}^*(10)$, Sv/s	Expanded uncertainty $U(\dot{H}^*(10))$, %	Personal dose equivalent rate $\dot{H}_p(10)$, Sv/s	Expanded uncertainty $U(\dot{H}_p(10))$, %
0.5	$1,456 \cdot 10^{-8}$	3.3	$2,534 \cdot 10^{-8}$	3.9	$2,752 \cdot 10^{-8}$	3.9
0.7	$7,365 \cdot 10^{-9}$	3.3	$1,282 \cdot 10^{-8}$	3.9	$1,392 \cdot 10^{-8}$	3.9
1.0	$3,615 \cdot 10^{-9}$	3.4	$6,290 \cdot 10^{-9}$	3.9	$6,833 \cdot 10^{-9}$	3.9
1.4	$1,846 \cdot 10^{-9}$	3.4	$3,212 \cdot 10^{-9}$	3.9	$3,489 \cdot 10^{-9}$	3.9
2.0	$8,758 \cdot 10^{-10}$	3.9	$1,524 \cdot 10^{-9}$	4.3	$1,655 \cdot 10^{-9}$	4.3

Signature of the person who has performed the calibration

Valiantsin Zubarau

(decryption of signature)

Continuation of the calibration protocol № 1-BY/04/2022 dated April 26, 2022

1.8. The actual value of the air kerma rate \dot{K}_A , Gy/s, at the position of the effective center of the ionization chamber (operating point) is determined by the formula

$$\dot{K}_A = N_{A0} \cdot M_0 \cdot k_{PT0} \cdot \prod_i k_{i0}$$

1.9. Typical uncertainty budget for assessing the accuracy of a measurement result:

Input value	Description	Meaning	Degrees of variance number	Relative standard uncertainty value u_r , %	
				type «A»	type «B»
N_{A0}	Calibration factor for the reference electrometer entered into the reference electrometer	Certificate	100	-	(1.3 - 1.5)
M_0	Electrometer readings minus electrometer readings when measuring current leakage and ambient radiation	Average electrometer readings	(6 - 15)	(0.00 - 0.98)	-
T_i	Ambient air temperature under measurement conditions	Thermometer readings	30	-	0.02
P_i	Atmospheric pressure at the level of the operating point	Barometer readings	30	-	0.006
k_{i0}	The difference between the temperature inside the ionization chamber and the thermometer readings	1	10	-	0.20
k_{20}	Long-term stability of the reference electrometer	1	10	-	0.17
k_{30}	Non-linearity of scales and mismatch of electrometer modes when switching modes	1	30	-	0.29
k_{40}	Resolution of the scale of indications of the reference electrometer	1	30	-	(0.00 - 0.03)
k_{50}	Transition from air kerma units to ambient or personal dose equivalent units	1	30	-	(0.0 - 1.0)
k_{60}	Orientation of the ionization chamber in the beam - uncertainty in the exact positioning of the reference and calibrated devices due to the inclination and (or) rotation of the ionization chamber relative to the measuring plane in the beam	1	10	-	0.10
k_{100}	The difference between the effective and geometric centers of the ionization chamber, in case the ionization chamber has a relatively large volume. Depends on the distance «source - ionization chamber» and the volume of the chamber	1	10	-	0.58
k_{120}	Difference in energy between the radiation used to calibrate the reference electrometer and the reference radiation used to calibrate	1	10	-	0.10
k_{130}	Beam inhomogeneity. A small ionization chamber determines the change in kerma power over the beam cross-sectional area	1	30	-	0.10
Standard uncertainty for types «A» and «B», %				(0.00 - 0.98)	(1.49 - 1.94)
Total standard uncertainty, %				(1.5 - 2.1)	
Degrees of variance effective number				(98 - 143)	
Expanded uncertainty, % ($k = 2$, $P = 95\%$)				(3.0 - 4.3)	
<i>The expanded uncertainty is obtained by multiplying the combined standard uncertainty by a coverage factor $k = 2$ corresponding to a confidence interval of approximately 95 % assuming a normal distribution. The evaluation of uncertainty is conducted according to the «Guide to the expression of uncertainty in measurement (GUM)».</i>					

Signature of the person who has performed the calibration

(signature)

Valiantsin Zubaraus

(decryption of signature)

ANNEX «A»

to the calibration protocol
(mandatory application)

Additional information

A1. Diameters of UDG-PM9000 calibration facility for gamma radiation beams, non-uniformity less than $\pm 3.0\%$:

- for collimator № 1 at 2.0 m distance from center of active part of a source to operating point – 0.60 m;
- for collimator № 2 at 1.0 m distance from center of active part of a source to operating point – 0.14 m;
- for collimator № 3 at 2.0 m distance from center of active part of a source to operating point – 0.60 m.

A2. Calculation of air kerma rates at the distances, which are not specified in the tables, is recommended to carry out taking into account the values for the closest distance, which is specified in the tables, and using the inverse square law. In this case it is also necessary to take into account air attenuation of gamma radiation. The recommended values for linear air attenuation coefficients at 20°C air temperature and 101.3 kPa air pressure, are:

- $22.577 \cdot 10^{-3} \text{ m}^{-1}$ (for ^{241}Am);
- $9.332 \cdot 10^{-3} \text{ m}^{-1}$ (for ^{137}Cs);
- $6.848 \cdot 10^{-3} \text{ m}^{-1}$ (for ^{60}Co).

A3. The recommended half-life periods of radionuclides:

- 1925.19 days (for ^{60}Co)*;
- 10975.5 days (for ^{137}Cs)*;
- 158004 days (for ^{241}Am)*.

* - reference data are given in the document «Monographie BIPM-5 – Table of radionuclides».

A4. Transition from air kerma rate to ambient dose equivalent rate $\dot{H}^*(10)$ and personal dose equivalent rate $\dot{H}_p(10)$ of gamma radiation is recommended to carry out using conversion coefficients specified in the document ISO 4037-3:2019:

$$\dot{H}^*(10) = h^*(10)_K \cdot \dot{K}_a$$

$$\dot{H}_p(10) = h_p(10)_K \cdot \dot{K}_a$$

where:

$$h^*(10)_K = 1.16 \text{ Sv/Gy} \quad \text{and} \quad h_p(10)_K = 1.15 \text{ Sv/Gy} \quad \text{- for } ^{60}\text{Co}$$

$$h^*(10)_K = 1.21 \text{ Sv/Gy} \quad \text{and} \quad h_p(10)_K = 1.21 \text{ Sv/Gy} \quad \text{- for } ^{137}\text{Cs}$$

$$h^*(10)_K = 1.74 \text{ Sv/Gy} \quad \text{and} \quad h_p(10)_K = 1.89 \text{ Sv/Gy} \quad \text{- for } ^{241}\text{Am}$$

Signature of the person who has performed the calibration

Valiantsin Zubara
(decryption of signature)

Signature of the person who has checked the calibration protocol

Sergey Sergeev
(decryption of signature)

End of protocol

-
- 1) Partial reproduction of the protocol is not allowed without the permission of Polimaster Ltd.
 - 2) Calibration results apply only to the calibration objects specified in the protocol.



Свидетельство о калибровке

Calibration certificate



БГЦА	BY/112.5.0100
BSCA	ГОСТ ISO/IEC 17025

Номер свидетельства BY 01 № 607 - 48 Дата калибровки по 08.10.2021 Страница 1 из 3
Certificate number Date when calibrated Page

Объект калибровки
Item calibrated

Дозиметр клинический серии PTW UNIDOS^{webline} type T10021 №000759 с ионизационными камерами TM32002 №000498, TM32003 №000163 и TM32005 №000097.

Наименование эталона / средства измерения / идентификация
Description of measurement standard / measuring instrument / identification

Заказчик
Customer

ООО «Полимастер», 220040, г. Минск, ул. Богдановича, 112-3и, каб. 53.

Информация о заказчике, адрес/ Name of the customer, address

Калибровочное клеймо-наклейка
Calibration mark

072943

порядковый номер
serial number

Метод калибровки

По мощности кермы в воздухе калибровка камер ионизационных выполнена в полях фотонного излучения от радионуклидных источников гамма-излучения ($E_{\gamma} = 59,54$ кэВ для ^{241}Am , $E_{\gamma} = 661,66$ кэВ для ^{137}Cs и $E_{\gamma} = 1252,87$ кэВ (средняя) для ^{60}Co) методом замещения в опорных точках эталонных ионизационных камер из состава Национального эталона кермы в воздухе и мощности кермы в воздухе НЭ РБ 7-01 в соответствии с методикой калибровки МРП МК 100055197.183-2021.

Method of calibration

Наименование метода / идентификация
Name of the method / identification

Все измерения имеют прослеживаемость к единицам Международной системы SI, которые воспроизводятся национальными эталонами НМИ. Данный сертификат может быть воспроизведен только полностью. Любая публикация или частичное воспроизведение содержания сертификата возможны с письменного разрешения БелГИМ, выдавшего сертификат.

All measurements are traceable to the SI units which are realized by national measurement standards of NMI.
This certificate shall not be reproduced, except in full. Any publication extracts from the calibration certificate requires written approval of the issuing BelGIM.



Начальник отдела


Ф.И.О

Кийко А.Н.

Дата выдачи

11.10.2021

Date of issue

Свидетельство о калибровке

Calibration certificate

с 06.10.2021

Номер свидетельства BY 01 № 607 - 48
Certificate number

Дата калибровки по 08.10.2021
Date when calibrated

Страница 2 из 3
Page

Калибровка выполнена с помощью
Calibration is performed by using

эталонного дозиметра клинического серии PTW UNIDOS^{webline} T10021 №000670 с ионизационными камерами TM32002 №0093 и Exradin A5 №XY150091, входящего в состав Национального эталона единицы кермы в воздухе и мощности кермы в воздухе НЭ РБ 7-01, имеющего прослеживаемость к Государственному первичному эталону единиц кермы в воздухе, мощности кермы в воздухе, экспозиционной дозы, мощности экспозиционной дозы и потока энергии рентгеновского и гамма-излучений ГЭТ 8-2019 Российской Федерации; эквивалентность НЭ РБ 7-01 подтверждена дополнительными сличениями COOMET №445/DE/08 (COOMET.RI(I)-S1 in KCDB).

Наименование эталонов и их статус / идентификация / доказательство прослеживаемости
Description of the reference measurement standards / identification / evidence of traceability

Условия калибровки
Calibration conditions

температура окружающей среды:
от 18,5 °C до 20,3 °C,
атмосферное давление:
от 100,5 до 101,4 кПа,
относительная влажность воздуха:
от 25,6 % до 47,4 %,
внешний радиационный фон:
не более 0,17 мкЗв/ч.

Условия окружающей среды и другие влияющие факторы
Environmental conditions and other influence parameters

Результаты калибровки,
включая неопределенность
Calibration results including uncertainty

Калибровочные коэффициенты и значения расширенной неопределенности для ионизационной камеры TM32002 №000498 (полярность, напряжение питания: **positive, 400 V**):

Качество излучения	Поддиапазон измерения	Энергия фотонного излучения E_{γ} , кэВ	Калибровочный коэффициент N_A , Гр/Кл	Расширенная неопределенность $U_I(N_A)$, %
S-Am	Low	59,54	$2,508 \cdot 10^4$	3,0
S-Cs	Low	661,66	$2,489 \cdot 10^4$	2,7
	Med		$2,500 \cdot 10^4$	2,7
S-Co	Low	1252,87	$2,460 \cdot 10^4$	2,7

Калибровочные коэффициенты и значения расширенной неопределенности для ионизационной камеры TM32003 №000163 (полярность, напряжение питания: **positive, 400 V**):

Качество излучения	Поддиапазон измерения	Энергия фотонного излучения E_{γ} , кэВ	Калибровочный коэффициент N_A , Гр/Кл	Расширенная неопределенность $U_I(N_A)$, %
S-Cs	Low	661,66	$3,085 \cdot 10^3$	2,7
	Med		$3,093 \cdot 10^3$	2,7
S-Co	Low	1252,87	$3,031 \cdot 10^3$	2,8

Калибровочные коэффициенты и значения расширенной неопределенности для ионизационной камеры TM32005 №000097 (полярность, напряжение питания: **positive, 400 V**):

Качество излучения	Поддиапазон измерения	Энергия фотонного излучения E_{γ} , кэВ	Калибровочный коэффициент N_A , Гр/Кл	Расширенная неопределенность $U_I(N_A)$, %
S-Cs	Low	661,66	$1,105 \cdot 10^6$	2,6
	Med		$1,101 \cdot 10^6$	2,6

Место проведения калибровки
Place of calibration

Старовиленский тракт, 93, 220053, г. Минск / Starovilensky trakt, 93, 220053, Minsk, Belarus
Тел /Phone.: 354-65-04; факс/ fax: 244-99-38 , e-Mail: info@belgim.by

Свидетельство о калибровке

Calibration certificate

с 06.10.2021

Номер свидетельства BY 01 № 607 - 48 Дата калибровки по 08.10.2021 Страница 3 из 3
Certificate number Date when calibrated Page

Расширенная неопределенность получена путем умножения суммарной стандартной неопределенности на коэффициент охвата $k = 2$, соответствующего уровню доверия приблизительно равному 95 % при допущении нормального распределения. Оценивание неопределенности проведено в соответствии с «Руководством по выражению неопределенности в измерениях (GUM)».

The expanded uncertainty is obtained by multiplying the combined standard uncertainty by a coverage factor $k = 2$ corresponding to a confidence interval of approximately 95 % assuming a normal distribution. The evaluation of uncertainty is conducted according to the «Guide to the expression of uncertainty in measurement (GUM)».

Дополнительная информация:

межкалибровочный интервал (рекомендуемый) –
не более 24 месяцев

Additional information

состояние объекта калибровки / регулировка и/или ремонт объекта калибровки до его калибровки
/condition of the item of calibration / adjustments or repair of the item of calibration before calibrated/

Подпись лица, выполнившего калибровку

Signature of the person who has performed calibration

Ведущий инженер,
Пархимчик А.О.

Должность, Ф.И.О \ Function, name

Подпись лица, выполнившего калибровку

Signature of the person who has performed calibration

Инженер 1-ой категории,
Климчена В.А.

Должность, Ф.И.О \ Function, name